

Certificate of Analysis

Product Compact Dry EC
for *E. coli* and coliforms
ID-No. 1 000 168 /1 000 169 /1 400 170
1 502 880
Lot Number 404007
New Expiry Date 30.06.2022

Test	Specifications	Results*
Appearance	Sheet of light yellow colour	Corresponds
Loss on drying	< 10, 0 %	Corresponds
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
pH value	6.8 -7.2	Corresponds
Colony growth		
<i>Escheria coli</i>	ATCC8739 Bluish green colonies	Corresponds
<i>Klebsiella oxytoca</i>	ATCC13182 Reddish purple colonies	Corresponds
<i>Pseudomonas aeruginosa</i>	ATCC9027 White colonies	Corresponds
<i>Bacillus subtilus</i>	ATCC6633 No growth	Corresponds
<i>Staphylococcus aureus</i>	ATCC6538 No growth	Corresponds

This lot meets the specifications.

Date: 14.12.2020

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.