

## Certificate of Analysis

**Product** Compact Dry EC  
for *E. coli* and coliforms  
**ID-No.** 1 000 168 /1 000 169 /1 400 170  
1 502 880  
**Lot Number** 409010  
**New Expiry Date** 30.09.2022

Test	Specifications	Results*
<b>Appearance</b>	Sheet of light yellow colour	Corresponds
<b>Loss on drying</b>	< 10, 0 %	Corresponds
<b>Aseptic test</b>	No growth of any bacteria. (30°C for 5 days. Visual check )	Corresponds
<b>pH value</b>	6.8 -7.2	Corresponds
<b>Colony growth</b>		
<i>Escheria coli</i>	ATCC8739 Bluish green colonies	Corresponds
<i>Klebsiella oxytoca</i>	ATCC13182 Reddish purple colonies	Corresponds
<i>Pseudomonas aeruginosa</i>	ATCC9027 White colonies	Corresponds
<i>Bacillus subtilus</i>	ATCC6633 No growth	Corresponds
<i>Staphylococcus aureus</i>	ATCC6538 No growth	Corresponds

**This lot meets the specifications.**

Date: 20.11.2020

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

\*Data is obtained either by OEM partner or HyServe.