

Certificate of Analysis

Product Compact Dry X-SA for *S. aureus*
ID-No. 1 002 960 / 1 002 961 / 1 002 962/
1 402 961 / 1 502 990 / 1 402 990
Lot Number 254012
Expiry Date 31.08.2022

Test	Specifications	Results*	
Appearance	Dry Sheet of light yellow colour	Corresponds	
Loss on drying	< 10,0 %	Corresponds	
Aseptic test	No growth of any bacteria. (30°C for 5 days, visual check)	Corresponds	
pH value	7.2 -7.6	Corresponds	
Colony growth			
<i>Staphylococcus aureus</i>	ATCC 6538	Growth, blue green	Corresponds
<i>Staphylococcus aureus</i>	ATCC 12600	Growth, blue green	Corresponds
<i>Staphylococcus epidermidis</i>	ATCC 14990	Growth, white	Corresponds
<i>Bacillus subtilis</i>	ATCC 6633	Growth, red	Corresponds
<i>Enterococcus faecalis</i>	ATCC 19433	No growth	Corresponds
<i>Escherichia coli</i>	ATCC 8739	No growth	Corresponds
<i>Pseudomonas aeruginosa</i>	ATCC 9027	No growth	Corresponds

This lot meets the specifications.

Date: 15.02.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.