

## Certificate of Analysis

**Product** Compact Dry X-SA for *S. aureus*  
**ID-No.** 1 002 960 / 1 002 961 / 1 002 962/  
1 402 961 / 1 502 990 / 1 402 990  
**Lot Number** 262106  
**Expiry Date** 28.02.2023

Test	Specifications	Results*	
<b>Appearance</b>	Dry Sheet of light yellow colour	Corresponds	
<b>Loss on drying</b>	< 10,0 %	Corresponds	
<b>Aseptic test</b>	No growth of any bacteria. (30°C for 5 days, visual check)	Corresponds	
<b>pH value</b>	7.2 -7.6	Corresponds	
<b>Colony growth</b>			
<i>Staphylococcus aureus</i>	ATCC 6538	Growth, blue green	Corresponds
<i>Staphylococcus aureus</i>	ATCC 12600	Growth, blue green	Corresponds
<i>Staphylococcus epidermidis</i>	ATCC 14990	Growth, white	Corresponds
<i>Bacillus subtilis</i>	ATCC 6633	Growth, red	Corresponds
<i>Enterococcus faecalis</i>	ATCC 19433	No growth	Corresponds
<i>Escherichia coli</i>	ATCC 8739	No growth	Corresponds
<i>Pseudomonas aeruginosa</i>	ATCC 9027	No growth	Corresponds

**This lot meets the specifications.**

Date: 19.05.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

\*Data is obtained either by OEM partner or HyServe.