

Certificate of Analysis

Product Compact Dry X-SA for *S. aureus*
ID-No. 1 002 960 / 1 002 961 / 1 002 962/
1 402 961 / 1 502 990 / 1 402 990
Lot Number 263106
Expiry Date 28.02.2023

| Test | Specifications | Results* | |
|-----------------------------------|---|--------------------|-------------|
| Appearance | Dry Sheet of light yellow colour | Corresponds | |
| Loss on drying | < 10,0 % | Corresponds | |
| Aseptic test | No growth of any bacteria. (30°C for 5 days, visual check) | Corresponds | |
| pH value | 7.2 -7.6 | Corresponds | |
| Colony growth | | | |
| <i>Staphylococcus aureus</i> | ATCC 6538 | Growth, blue green | Corresponds |
| <i>Staphylococcus aureus</i> | ATCC 12600 | Growth, blue green | Corresponds |
| <i>Staphylococcus epidermidis</i> | ATCC 14990 | Growth, white | Corresponds |
| <i>Bacillus subtilis</i> | ATCC 6633 | Growth, red | Corresponds |
| <i>Enterococcus faecalis</i> | ATCC 19433 | No growth | Corresponds |
| <i>Escherichia coli</i> | ATCC 8739 | No growth | Corresponds |
| <i>Pseudomonas aeruginosa</i> | ATCC 9027 | No growth | Corresponds |

This lot meets the specifications.

Date: 19.09.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.