

CERTIFICATE OF ANALYSIS**Easy Plate AC**

Product code :61980

Lot No.: AC20210913A

Expiry date: Mar. 12, 2023

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	*See below
Spreading time	passed test	≤1min
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≤4mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

*The ratio of the average colony count on the product to the standard agar medium(%).

Escherichia coli (NBRC15034) : 70-200

Bacillus subtilis (NBRC3134) : 70-350

Klebsiella pneumouniae (NBRC14940) : 50-200

Pseudomonas aeruginosa (NBRC3899) : 70-350

Staphylococcus aureus (ATCC25923) : 50-200

Enterococcus faecalis (NBRC100482) : 70-200

Storage condition : 2-8°C


Hiroshi Katayama; Manager

Date Dec. 21, 2021

KBG6198000aa-Eaa
SA2112-0213

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company