

**CERTIFICATE OF ANALYSIS****Easy Plate EC**

Product code :61982

Lot No.: EC20210803A

Expiry date: Feb. 02, 2023

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	* See below
Spreading time	passed test	$\leq$ 1min
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	$\leq$ 4mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

\*The ratio of the average colony count on the product to the VRB agar medium(%).

Escherichia coli (NBRC15034) : 70-200  
Klebsiella pneumouniae(NBRC14940) : 50-200  
Enterobacter cloacae (NBRC13536) : 50-200  
Citrobacter freundii (NBRC12681) : 50-200  
Staphylococcus aureus (ATCC25923) : No growth  
Enterococcus faecalis (NBRC100482) : No growth  
Proteus mirabilis (NBRC105697) : No growth  
Staphylococcus xylosus (NBRC109770) : No growth

Storage condition : 2-8°C

  
Ryosuke Sugimoto; Manager

Date Oct. 19, 2021KBG6198200ab-Eaa  
SA2110-0156Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company