

CERTIFICATE OF ANALYSIS**Easy Plate SA**

Product code :61983

Lot No.: SA20210806A

Expiry date: Aug. 05, 2022

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	* See below
Spreading time	passed test	≤1min
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≤4mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

* The ratio of the average colony count on the product to the standard agar medium (%).

Staphylococcus aureus (ATCC25923) : 50-200 (as blue colony)

Staphylococcus aureus (NBRC13276) : 50-200 (as blue colony)

Enterococcus faecalis (NBRC100482) : No growth

Bacillus subtilis (NBRC3134) : No growth

Escherichia coli (NBRC15034) : No growth

Proteus mirabilis (NBRC105697) : No growth

Bacillus licheniformis (NBRC12200) : Growth as pink colony

Staphylococcus epidermidis (NBRC100911) : ≤10 and no growth as blue colony

Storage condition : 2-8°C


Ryosuke Sugimoto; Manager

Date Oct. 19, 2021KBG6198300aa-Eaa
SA2110-0157Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company