

CERTIFICATE OF ANALYSIS**Easy Plate SA**

Product code :61976

Lot No.: SA20220721A

Expiry date: Jul. 20, 2023

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	≤ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≤ 4 mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

Performance test

Staphylococcus aureus (ATCC25923) : 50-200 (as blue colony)^{*1}Staphylococcus aureus (NBRC13276) : 50-200 (as blue colony)^{*1}

Enterococcus faecalis (NBRC100482) : No growth

Bacillus subtilis (NBRC3134) : No growth

Escherichia coli (NBRC15034) : No growth

Proteus mirabilis (NBRC105697) : No growth

Bacillus licheniformis (NBRC12200) : Growth as pink colony

Staphylococcus epidermidis (NBRC100911) : ≤ 10 and no growth as blue colony^{*2}^{*1}The ratio of the average colony count on the product to the plate count agar (%).^{*2}The ratio of the average colony count on the product to Easy Plate AC (%).

Storage condition : 2-8°C


Ryosuke Sugimoto; ManagerDate Sep. 21, 2022KBG6197600aa-Eab
SA2209-0212Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company