Certificate of Analysis

HyServe
GmbH & Co.KG.

Hechenrainer Str 12
82449 Uffing
Ph.+49(0)8846-1344
Fax+49(0)8846-1342

info@hyserve.com
www.hyserve.com

Product	**Listeria Swab**
ID-No.	**1 000 100 / 1 000 100L**
Lot Number	**21K26**
Expiry Date	**31.10.2022**

Test		**Specifications**	**Results***
Appearance		Light brown colour	Corresponds
Aseptic test		No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
Colony growth			
Listeria monocytogenes	NCTC 5214	Black colour development	Corresponds
Listeria innocua	NCTC 11288	Black colour development	Corresponds
Listeria monocytogenes	NCTC 11994	Black colour development	Corresponds
Listeria monocytogenes	7973	Black colour development	Corresponds
Staphylococcus aureus	*NCTC 8532*	No colour development	Corresponds
Staphylococcus faealis	*NCTC 0775*	No colour development	Corresponds
Lactobacillus spec		No colour development	Corresponds
Listeria grayli	*NCTC 10815*	No colour development	Corresponds

This lot meets the specifications.

Date: 02.12.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.