

**CERTIFICATE OF ANALYSIS****LuciPac A3 Surface**

Product code :60361

Lot No.: 20230411Y

Expiry date: Jul. 10, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 $\mu$ L of 5pmol ADP Average Luminescence(RLU)	13100	$\geq 9100, \leq 16900$
When applied 100 $\mu$ L of 5pmol AMP Average Luminescence(RLU)	13100	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
\_\_\_\_\_  
Ryosuke Sugimoto; ManagerDate May. 11, 2023KBG6036100ac-Eaa  
SA2305-0106Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company