



**CERTIFICATE OF ANALYSIS**

**LuciPac A3 Surface**

Product code :60361

Lot No.: 20210402Y

Expiry date: Jul. 01, 2022

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 $\mu$ L of 5pmol ADP Average Luminescence(RLU)	12600	$\geq 9100, \leq 16900$
When applied 100 $\mu$ L of 5pmol AMP Average Luminescence(RLU)	12600	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

*R. Sugimoto*  
Ryosuke Sugimoto; Manager

Date Apr. 22, 2021

KBG6036100ac-Eaa  
SA2104-0159

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company