

CERTIFICATE OF ANALYSISLuciPac A3 Surface

Product code :60361


Lot No.: 20230118Y

Expiry date: Apr. 17, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 $\mu$ L of 5pmol ADP Average Luminescence(RLU)	13500	$\geq 9100, \leq 16900$
When applied 100 $\mu$ L of 5pmol AMP Average Luminescence(RLU)	13500	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
\_\_\_\_\_  
Ryosuke Sugimoto; ManagerDate Mar. 07, 2023KBG6036100ac-Eaa  
SA2303-0084Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company