



CERTIFICATE OF ANALYSIS

LuciPac A3 Water

Product code :60365


Lot No.: 20210430W

Expiry date: Jul. 29, 2022

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	13500	$\geq 9100, \leq 16900$
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	13300	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Ryosuke Sugimoto; Manager

Date May. 19, 2021

KBG6036500ac-Eaa
SA2105-0139

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company