



CERTIFICATE OF ANALYSIS

LuciPac Pen

Product code :60331

Lot No.: 20211102U

Expiry date: Feb. 01, 2023

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12900	$\geq 9100, \leq 16900$
(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)		

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Ryosuke Sugimoto; Manager

Date Nov. 11, 2021

KBG6033100ab-Eaa
SA2111-0075

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company