



CERTIFICATE OF ANALYSIS

LuciPac Pen

Product code :60331

Lot No.: 20211115U

Expiry date: Feb. 14, 2023

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12800	$\geq 9100, \leq 16900$
(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)		

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Ryosuke Sugimoto; Manager

Date Dec. 06, 2021

KBG6033100ab-Eaa
SA2112-0032

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company