



**CERTIFICATE OF ANALYSIS**

**LuciPac Pen**

Product code :60331

---

Lot No.: 20211124U

Expiry date: Feb. 23, 2023

---

| <u>TEST PARAMETER</u>  | <u>RESULT</u> | <u>SPECIFICATION</u>    |
|--|---------------|-------------------------|
| Average Luminescence(RLU)  | 12300         | $\geq 9100, \leq 16900$ |
| (When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart) |               |                         |

---

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
\_\_\_\_\_  
Ryosuke Sugimoto; Manager

Date Dec. 06, 2021

KBG6033100ab-Eaa  
SA2112-0033

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company