

CERTIFICATE OF ANALYSIS**LuciPac Pen**

Product code :60331

Lot No.: 20211201U

Expiry date: Feb. 28, 2023

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Hiroshi Katayama; Manager

Date Dec. 21, 2021KBG6033100ab-Eaa
SA2112-0206Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company