



CERTIFICATE OF ANALYSIS

LuciPac Pen-AQUA

Product code :60336


Lot No.: 20210621Q

Expiry date: Sep. 20, 2022

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12600	$\geq 9100, \leq 16900$

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; Manager

Date Jul. 14, 2021

KBG6033600ab-Eaa
SA2107-0130

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company