



**CERTIFICATE OF ANALYSIS**

**LuciPac Pen-AQUA**

Product code :60336

---

Lot No.: 20211001Q

Expiry date: Dec. 31, 2022


---

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12500	$\geq 9100, \leq 16900$

(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)

---

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
\_\_\_\_\_  
Ryosuke Sugimoto, Manager

Date Nov. 11, 2021

KBG6033600ab-Eaa  
SA2111-0077

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company