

CERTIFICATE OF ANALYSIS**LuciPac Pen-AQUA**

Product code :60336

Lot No.: 20221003Q

Expiry date: Jan. 02, 2024

| <u>TEST PARAMETER</u> | <u>RESULT</u> | <u>SPECIFICATION</u> |
|---------------------------|---------------|----------------------------|
| Average Luminescence(RLU) | 12700 | ≥ 9100 , ≤ 16900 |

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Ryosuke Sugimoto; ManagerDate Dec. 22, 2022KBG6033600ab-Eaa
SA2212-0218Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company