



Kikkoman Biochemifa Company

CERTIFICATE OF ANALYSIS

LuciPac Pen-AQUA

Product code :60336

Lot No.: 20221209Q

Expiry date: Mar. 08, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	13200	$\geq 9100, \leq 16900$

(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Ryosuke Sugimoto; Manager

Date Mar. 07, 2023

KBG6033600ab-Eaa  
SA2303-0093

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company