



CERTIFICATE OF ANALYSIS

LuciPac Pen

Product code :60331

---

Lot No.: 20210308U

Expiry date: Jun. 07, 2022


---

| <u>TEST PARAMETER</u>     | <u>RESULT</u> | <u>SPECIFICATION</u>    |
|---------------------------|---------------|-------------------------|
| Average Luminescence(RLU) | 12800         | $\geq 9100, \leq 16900$ |

(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)

---

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
\_\_\_\_\_  
Ryosuke Sugimoto; Manager

Date Mar. 18, 2021

KBG6033100ab-Eaa  
SA2103-0175

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company