



**CERTIFICATE OF ANALYSIS**

**LuciPac Pen**

Product code :60331

---

Lot No.: 20210310U

Expiry date: Jun. 09, 2022


---

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12900	$\geq 9100, \leq 16900$

(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)

---

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
\_\_\_\_\_  
Ryosuke Sugimoto; Manager

Date Apr. 22, 2021

KBG6033100ab-Eaa  
SA2104-0165

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company