



CERTIFICATE OF ANALYSIS

LuciPac Pen

Product code :60331


Lot No.: 20210406U

Expiry date: Jul. 05, 2022

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	11100	$\geq 9100, \leq 16900$

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Ryosuke Sugimoto; Manager

Date Apr. 22, 2021

KBG6033100ab-Eaa
SA2104-0156

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company