

CERTIFICATE OF ANALYSISLuciPac Pen

Product code :60331


Lot No.: 20210705U

Expiry date: Oct. 04, 2022

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12600	$\geq 9100, \leq 16900$

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; ManagerDate Jul. 14, 2021KBG6033100ab-Eaa
SA2107-0126Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company