



CERTIFICATE OF ANALYSIS

LuciPac Pen

Product code :60331

Lot No.: 20210708U

Expiry date: Oct. 07, 2022

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
-----------------------	---------------	----------------------

Average Luminescence(RLU)	12500	$\geq 9100, \leq 16900$
---------------------------	-------	-------------------------

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Ryosuke Sugimoto; Manager

Date Jul. 14, 2021

KBG6033100ab-Eaa
SA2107-0129

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company