



CERTIFICATE OF ANALYSIS

LuciPac Pen

Product code :60331


Lot No.: 20220324U

Expiry date: Jun. 23, 2023

| <u>TEST PARAMETER</u> | <u>RESULT</u> | <u>SPECIFICATION</u> |
|---------------------------|---------------|-------------------------|
| Average Luminescence(RLU) | 12500 | $\geq 9100, \leq 16900$ |

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; Manager

Date Apr. 04, 2022

KBG6033100ab-Eaa
SA2204-0028

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company