

CERTIFICATE OF ANALYSISLuciPac Pen

Product code :60331

Lot No.: 20220412U

Expiry date: Jul. 11, 2023

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$

(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
\_\_\_\_\_  
Ryosuke Sugimoto; ManagerDate May. 25, 2022KBG6033100ab-Eaa  
SA2205-0196Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company