



Kikkoman Biochemifa Company

CERTIFICATE OF ANALYSIS

LuciPac Pen

Product code :60331

Lot No.: 20220609U

Expiry date: Sep. 08, 2023

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12600	$\geq 9100, \leq 16900$

(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Ryosuke Sugimoto; Manager

Date Jul. 19, 2022

KBG6033100ab-Eaa  
SA2207-0146

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company