



Kikkoman Biochemifa Company

CERTIFICATE OF ANALYSIS

LuciPac Pen

Product code :60331

Lot No.: 20220705U

Expiry date: Oct. 04, 2023

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12500	$\geq 9100, \leq 16900$
(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)		

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Ryosuke Sugimoto; Manager

Date Jul. 19, 2022

KBG6033100ab-Eaa
SA2207-0148

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company