



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

November 14, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.
Quality Assurance Department
Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	22NOSFSY062	NOV.12 2023

The product kit, batch 22NOSFSY062, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : November 14, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.011	0.009	0.011	0.010	0.010	
0.78	0.051	0.051	0.053	0.051	0.052	1.9
1.56	0.092	0.095	0.097	0.095	0.095	2.2
3.13	0.175	0.174	0.177	0.174	0.175	0.8
6.25	0.338	0.341	0.336	0.329	0.336	1.5
12.5	0.640	0.635	0.638	0.632	0.636	0.6
25	1.233	1.238	1.228	1.209	1.227	1.0
50	2.173	2.189	2.208	2.219	2.197	0.9

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.7	6.2-10.1
B	9.6	8.5-13.8
C	5.7	4.4-7.8